

1/1B, Choudhary Hetram House, Bharat Nagar, New Friends Colony, New Delhi - 110 025 M:(+91) 9958994797, 8800257774, Ph:011-23557517; Email: procurement@samsconsult.com

RFQ No. SAMS/FIND/PCR Workstation/PROCT/LT/25/2015

Dated: 21/08/2015

# Limited Tender for Procurement of PCR Workstation, UV Light AND Desktop Deep freeze storage box with lid for 0.2ml PCR tubes

## **REQUEST FOR QUOTATIONS (RFQ)**

To,

at Annexure-III.

| Dear Sir,                               |  |
|---|--|
|   | Services Pvt. Ltd. (SAMS) has been engaged by Foundation for   |
| • | ia for providing procurement consultancy services for equipment, B Laboratories across India under GFATM Project. Accordingly, |

The SAMS hereby request you to submit your Quotations giving in full specifications, name of the manufacturer, catalogue etc. as per our terms and conditions enclosed herewith, latest by **28<sup>th</sup> August**, **2015 till 2:30 PM**.

SAMS on behalf of FIND is pleased to invite Quotation for Supply of PCR- Workstations, UV light and Desktop Deep freeze storage box with lid for 0.2ml PCR tubes, as per detailed specifications enclosed at Annexure-I, on CIP-Final Destinations terms of delivery basis, as per schedule of requirement enclosed

| Sch. No. | Particulars of item                      | Unit | Quantity |
|----------|--|------|----------|
|          | PCR-Workstation, UV light                | Nos. | 6        |
| I        | Desktop Deep freeze storage box with lid | Nos. | 70       |
|          | for 0.2ml PCR tubes                      |      |          |

The quotations with at least 3 months of validity may be sent in double cover duly wax sealed and prominently subscribed "RFQ No. SAMS/FIND/PCR Workstation/PROCT/LT/25/2015" to be opened on 28<sup>th</sup> August, 2015 at 03.00 PM.

The quotations, which are not received duly sealed, mentioning this office letter number and last date of receiving will not be accepted. The firms must ensure that besides all these terms and conditions enclosed herewith, the following documents must be submitted in the chronological order (strict compliance) to show their credentials against required qualifications and to avoid rejection of tenders.

- i. Forwarding letter clearly mentioning the authorized signatory, total number of pages, the model quoted, quotation validity undertaking of 3 months, which can be extended on mutual consent.
- ii. The instruments such as power of attorney, resolution of board etc., authorizing an officer of the bidder and verifying his signature, duly signed by the Authorized signatory of the company/firm. Only such authorized officer of the bidder should sign the tender documents.
- iii. Manufacturer's authorization form as per format attached (Annexure IV) with RFQ/ certificate of authorized dealership.
- iv. Copy of the incorporation / registration certificate issued by an appropriate authority
- v. Technical compliance statement of quoted equipment against required specifications of RFQ, as per format enclosed under technical specifications at Annexure-I.
- vi. Warranty for 3 years from the date of Installation must be given (No conditional warranty shall be accepted).
- vii. Technical Catalogue / brochure of item quoted as per proposed specifications.
- viii. Documents regarding quality standard (example ISO, BIS, CE, US FDA, electrical safety standards etc.) as per the quality criterion specified under Technical Specifications.
- ix. Details (address and contact details) of service centers/ branch offices available in India.
- x. Copies of Purchase Orders issued to your firm, showing experience of the supply and after sale service support for at least two similar contracts for the items included in this RFQ over the period of last three years.
- xi. At least two Client's Certificate for satisfactory / timely supply and installation of similar equipment, during last three years.
- xii. Annual turnover statement for last 3 financial years (2012-13, 2013-14, 2014-15) duly certified by the Chartered Accountant/ Income tax return certificates/ financial statement.
- xiii. VAT/ Excise/Sales tax/Income Tax registration and clearance certificate.

It should be addressed and submitted to the "STRATEGIC ALLIANCE MANAGEMENT SERVICES PVT. LTD. (SAMS), 1/1 B, Choudhary Hetram House, Bharat Nagar, New Friends Colony, New Delhi 110025, INDIA" and should reach this office on or before 28<sup>th</sup>August, 2015 till 02:30 PM.

Yours sincerely.

For Strategic Alliance Management Services Pvt. Ltd.

(Sanjay Rastogi)
Associate Director (MCS)

#### RFQ No. SAMS/FIND/PCR Workstation/PROCT/LT/25/2015

## TERMS & CONDITIONS AND INSTRUCTIONS TO THE BIDDER (ITB)

- Sealed tender superscribed "Limited Tender for procurement of PCR Workstation, UV Light and Desktop Deep freeze storage box with lid for 0.2ml PCR tubes" are invited for the detailed in the Schedule (attached herewith) to the "STRATEGIC ALLIANCE MANAGEMENT SERVICES PVT. LTD. (SAMS), 1/1 B, Choudhary Hetram House, Bharat Nagar, New Friends Colony, New Delhi 110025, INDIA".
- 2. The bidder is expected to examine all instructions, forms, terms and specifications in the bidding documents. Failure to furnish all information required by the bidding documents or submission of a bid not substantially responsive to the bidding documents in every respect will be at the bidder's risk and may result in the rejection of its bid.
- 3. The tender is to be submitted alongwith duly signed Terms & Conditions & ITB in a sealed cover which must be clearly marked with the "Limited Tender for procurement of PCR Workstation, UV Light and Desktop Deep freeze storage box with lid for 0.2ml PCR tubes" and the due date for its opening. The cover should be addressed to STRATEGIC ALLIANCE MANAGEMENT SERVICES PVT. LTD. (SAMS), 1/1 B, Choudhary Hetram House, Bharat Nagar, New Friends Colony, New Delhi 110025, INDIA.
- **4.** The bid shall contain no interlineations. erasures or overwriting except as necessary to correct errors made by the bidder, in which case such corrections shall be initiated by the person or persons signing the bid.
- **5.** Each bidder is entitled to submit only one tender wherein he can, if necessary indicate alternative quotation. In case where more than one tender in a specified group is submitted by him will be liable to rejection.
- **6.** In the event of the space on the prescribed form being insufficient for the required purposes, additional page must be numbers of consecutively bear the tender number and be fully signed by the bidder. In such cases reference to the additional pages must be made in the tender form.
- 7. The tender must reach this office strictly not later than 28<sup>th</sup> August, 2015 till 02:30 PM & the bid must be received by the address specified & no later than the time and date specified in the invitation forbids. In the event of the specified date for the submission of bids being declared a holiday for the purchaser, the bids will be received up to the appointed time on the next working day.
- **8.** It is the responsibility of the bidders to see that the complete bidding documents, whether sent by post or by courier or by person are received by the time and date stipulated for receipt failing which the bid would be considered late and rejected.
- **9.** Any bid received after the deadline for submission of bids prescribed will be rejected and/or returned unopened to the bidder.
- **10.** The bidder may modify or withdraw its bid after the bid's submission provided that written notice of the modification or withdrawal is received by the purchaser prior to the deadline prescribed for submission or bids.
- **11.** No bid may be modified subsequent to the deadline for submission of bids.

- **12.** All supplies should be subject to inspection on receipt and purchaser shall have full discretion to accept or reject. Rejected supply should be removed within 15 days from date of intimation otherwise it will be disposed off in any manner deemed fit by the SAMS/FIND/ Consignees.
- **13.** If the firm fails to supply the items ordered within the stipulated period. The firm will be debarred from SAMS approved vendor list for next three Years.
- **14.** The conditions of the bidder shall not be binding to the SAMS.
- **15.** Delivery schedule –Delivery/ installation should be completed before 30<sup>th</sup>September 2015. Bidder should clearly mention about the delivery period, if supplier is unable to supply the goods before stipulated time (before 30<sup>th</sup> September, 2015) will be treated as non-responsive. Delivery of goods to be carried out as per the **Consignee Distribution List** (Enclosed at **Annexure III**).

## 16. Terms of Delivery-

CIP- final place of destination as per Consignee Distribution List.

- a. The responsibility of arranging all required documents for supply of goods till final place of destinations, including Road Permits, unloading of equipment and shifting inside laboratory etc. is of the Supplier. The SAMS may only facilitate the Road Permits/ declarations for the same.
- b. The payment of any kind of taxes, octroi, entry tax, local taxes etc. shall be the responsibility of supplier and should be included in the financial quote.
- c. Installation of Equipment as per the Consignee Distribution List is the responsibility of supplier.
- d. Training & Commissioning of Equipment as per the Consignee Distribution List is the responsibility of supplier.

## 17. BID PRICES:

- a. The bidder shall indicate on the appropriate Price Form attached to these, the unit prices and total Quotation/tender prices of goods it proposes to supply under the contract. Price should be quoted only in INR in the specified format attached in **Annexure II**.
- b. Total Price till final Place of Destination and installation, as quoted by the bidders should be inclusive of all applicable taxes and duties (Excise Duty, Custom Duty, Entry tax, Road Tax, Octroi Duty and other local taxes etc.), only VAT/ CST to be quoted separately, as applicable. It should also include that all goods shall be duly insured up to the final place of destination.

#### 18. Bid currency

The prices should be quoted only in Indian National Rupees (INR).

#### 19. Document establishing bidder's eligibility and qualifications:

The documentary evidence of the bidder's qualifications to perform the contract if its bid/tender is accepted, shall establish to the purchaser's satisfaction:

- a. That, in the case of bidder offering to supply goods under the contract which the bidder did not manufacture or otherwise produce, the bidder has been duly authorized (as per authorization form in **Annexure-IV**) by the good's manufacturer or producer to supply the goods in India.
- b. That, in the case of a bidder not doing business within India, the bidder is or will be represented by an agent in India equipped and able to carry out the supplier's maintenance; repair and spare parts stocking obligations prescribed by the conditions of contract and/or Technical Specifications

- c. The bidder shall furnish, as part of its bid, documents establishing the bidder's eligibility to bid/tender and its qualifications to perform the contract if its bid/tender is accepted.
- d. The documentary evidence of the bidder's eligibility to bid/tender shall establish to the purchaser's action that the bidder, at the time of submission of its bid, is eligible.

## e. Financial Capability

The Bidder shall furnish documentary evidence that it meets the following financial requirement(s):

- i. The Bidders who are manufacturers should have annual sales turnover of minimum 3 times the estimated value for each schedule during any of the preceding three years, to qualify for a particular schedule. The above criteria will be cumulative for the purpose of evaluating multiple schedules.
- **ii.** For non manufacturer Bidders (excluding agents quoting in the name of the manufacturer): the Bidder should have annual sales turnover of minimum 2 times the estimated value for each schedule, during any of the preceding three years, to qualify for a particular schedule. The above criteria will be cumulative for the purpose of evaluating multiple schedules.

## 20. Period of validity of bids:

- a. Bids shall remain valid for 3 months after the date of bid/Tender opening prescribed by the purchaser, A bid/tender valid for a shorter period shall be rejected by the purchaser as non-responsive.
- b. In exceptional circumstances, the purchaser may solicit the bidder's consent to an extension of the period of bid/tender validity. The request and the responses thereto shall be made in writing (or by cable or by telex).

## 21. Format and signing of bid

- a. The bidder shall prepare two copies of the bid, clearly making each "Original Bid" and "Copy of Bid" as appropriate. In the event of any discrepancy between them, the original shall govern.
- b. The original and all copies of the bid/tender shall be typed or written in indelible ink and shall be signed by the bidder or a person duly authorized to bind the bidder to the contract. The letter authorization shall be indicated by written power-of-attorney accompanying the bid. All pages of the bid, except for unammended printed literature, shall be initialed by the person or person's signing the bid.
- c. The bid/tender shall contain no interlineration, erasures or overwriting except as necessary to correct errors made by the bidder, in which case such corrections shall be initialed by the person or persons signing the bid.

## 22. Preliminary examination

- a. The purchaser will examine the bids to determine whether they are complete, whether any computational errors have been made, whether the documents have been properly signed, whether the bid/ tender validity is as required and whether the bids are generally, in order.
- b. Bids from agents without proper authorization from the manufacturers or on principal's original Performa invoice with break-up of price shall be treated as non-responsive and shall be rejected.
- c. Arithmetical errors will be rectified on the following basis. If there is a discrepancy between the unit price and the total price that is obtained by multiplying the unit price and quantify, the unit price shall prevail and the total price shall be corrected. If the supplier does not accept the correction of

the errors, its bid/ tender will be rejected. If there is a discrepancy between the words and figures, the amount in. words shall prevail.

- d. Prior to the detailed evaluation purchaser will determine the substantial responsive-ness of each bid/tender to the bidding documents. For purpose of these clauses, a substantially responsive bid/tender is one which, conforms to all the terms and conditions of the bidding documents without material deviations. The purchaser's determination of a bid's responsiveness is to base on the contents of the bid/tender itself without recourse to extrinsic evidences.
- e. A bid/tender determined as not substantially responsive will be rejected by the purchaser and may not subsequently be made responsive by the bidder by correction of the non-conformity.
- f. The purchaser may waive any minor informality or non-conformity or irregularity in a bid, which does not constitute a material deviation, provided such waiver does not prejudice and affect the relative ranking of any bidder.

## 23. Evaluation and comparison of bid

- a. The purchaser will technically evaluate all bids previously determined to be responsive and compare the bids previously determined to be substantially responsive and are acceptable.
- b. The comparison of price shall be considering, delivery of goods as per CIP-till final place of destinations, installation, commissioning and training at consignee's end.
- c. The Purchaser's evaluation of a bid will take into account the total unit cost of the item at the consignee's destination inclusive of all kind of duties, taxes and other charges.
- d. The contract shall be awarded only to the bidder who are substantially responsive, offer competitive rates, and meet the qualification requirement stipulated in the bidding documents.
- e. Bidder may bid for one or more schedules (if applicable). Bids will be evaluated for each schedule separately and the contract will comprise the schedules(s) awarded to the successful bidder. Bidders must quote for the entire quantity of each schedule. Bidders who do not quote for full quantity of the schedule will be treated as non-responsive.
- f. Deviations in the delivery schedule and Payment schedule are not permitted.
- g. The Purchaser shall compare all substantially responsive bids to determine the lowest-evaluated bid, in accordance with evaluation criteria mentioned in this bidding document.

## 24. Award of Contract (Pre-qualification)

- a. Notwithstanding the 'Qualification Requirements' set out in terms & condition the purchaser will determine to the satisfaction whether the bidder selected as having submitted the lowest evaluated responsive bid/tender is qualified to satisfactorily perform the contract.
- b. The determination will take into account the bidder's financial, technical and production capabilities. It will be based upon the examination of the documentary evidence of the bidder's qualifications submitted by the bidder, as well as such other information as the purchaser deems necessary and appropriate.
- c. An affirmative determination will be a prerequisite for award of the contract to the bidder. A negative determination will result in rejection of the bidder's bid/tender in which event the purchaser will proceed to the next lowest evaluated bid/tender to make a similar determination of the bidder's capabilities to perform satisfactorily.

## d. Technical Demonstration of equipment

Prior to contract award, the lowest evaluated substantially responsive Bidder will be requested to organize demonstration of the equipment, for inspection of quality and its efficient operation and performance before SAMS / FIND or its authorized team. In case, the equipment does not perform satisfactorily, the bid would be deemed to be non-responsive and SAMS reserves a right to move to next lowest evaluated responsive bidder.

## 25. Award criteria

Subject to clause 24, the purchaser will award the contract to the successful bidder/ bidder whose bid/tender has been determined to be substantially responsive and has been determined as the lowest evaluated bid, provided further that the bidder/bidder is determined to be qualified to perform the contract satisfactorily.

## 26. Purchaser's right to vary quantities at time of award

The purchaser reserves the right at the time of award to increase or decrease up to 25% of the quantity of goods and services specified in the 'Schedule oh requirements' without any change in prices or other terms and conditions.

## 27. Purchaser's right to accept any bid/tender and to reject any or all bids

The purchaser reserves the right to accept or reject any bid/tender and to annul the bidding process and reject any or all bids at any time prior to award of contract, without thereby incurring any liability to the affected bidder/bidder or bidders or any obligations to inform the affected bidder/bidder of the grounds for the purchaser's action.

## 28. Performance Security

- a. Within twenty one (21) days from date of the issue of Purchase Order by the purchaser, the supplier, shall furnish performance security to the purchaser for an amount equal to **five percent** (5%) of the total value of the contract, valid for forty months from the date of Purchase Order.
- b. The Performance security shall be denominated in Indian Rupees or in the currency of the contract.
- c. It shall be in any one of the forms namely Account Payee Demand Draft or Fixed Deposit Receipt drawn from any Scheduled commercial bank in India or Bank Guarantee issued by a Scheduled commercial bank in India, in the prescribed form as provided in Annexure V of this document in favour of the purchaser. The validity of the Fixed Deposit receipt or Bank Guarantee will be for a period for forty months from the date of Purchase Order.
- d. In the event of any loss due to supplier's failure to fulfill its obligations in terms of the contract, the amount of the performance security shall be payable to the purchaser to compensate the purchaser for the same.
- e. In the event of any amendment issued to the contract, the supplier shall, within twenty-one (21) days of issue of the amendment, furnish the corresponding amendment to the Performance Security (as necessary), rendering the same valid in all respects in terms of the contract, as amended.
- f. Subject to sub clause d above, the purchaser will release the Performance Security without any interest to the supplier on completion of the supplier's all contractual obligations including the warranty obligations.

## 29. Spare parts

As specified in the contract Form, the supplier may be required to provide any or all of the following materials and notifications pertaining to spare parts manufactured and/or distributed by the supplier.

- a. Such spare parts as the purchaser may elect to purchase from the supplier, provided that this election will not relieve the supplier of any warranty obligations under the contract.
- b. In the event of termination of production of the spare parts advance notification to the purchaser of the pending termination, in sufficient time to permit the purchaser to procure needed requirements and following such termination, furnishing at no cost to the purchaser, the blueprints, drawings and specifications of the spare parts, if and when requested.
- c. Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spares such as gaskets, plugs, washers, belts, etc. Other spare parts and components shall be supplied as promptly as possible but in any case within six months of placement of order and establishment of letter of credit.

#### 30. Warranty

- a. The supplier warrants that the goods supplied under the contract are new, unused, of the most recent of current models and incorporate all recent improvements in design and materials unless provided otherwise in the contract. The supplier further warrants that the goods supplied under the contract shall have no defect arising from design, materials or -workmanship (except insofar as the design or material is required by the purchaser is specifications) or from any act or omission of the supplier that may develop under normal use of the supplied goods in India, i.e. the counter of final destination.
- b. The warranty shall remain valid for **36 (Thirty six) months** on the equipment or any portion thereof as the case may have been delivered to the final destination and installed, commissioned and take over by the consignee to the entire satisfaction of, the purchaser.
- c. The purchaser/indenter shall promptly notify the supplier in writing of any claim arising under this warranty.
- d. Upon receipt of such notice, the supplier shall, with all reasonable speed, repair or replace the defective goods or parts thereof, free of cost at the ultimate destination. The supplier shall elicit over the replaced parts/goods at the time of their replacement. No claim, whatsoever shall lie on the purchaser for the replaced parts/goods thereafter.
- e. If the supplier, having been notified, fails to remedy the defects) within a reasonable period, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expense and without prejudice to any other rights which the purchase may have against the supplier under the contract.
- f. The warranty for defective parts will begin de novo from the date of replacement. Supplier will pay customs duty and all expenses up to the destination for the replaced parts.

## 31. Payments Terms

The method and conditions of payment to be made to the Supplier (Payments will not be made to any other party) under this tender, as applicable, shall be as follows:

- i) On Delivery to Consignee: Ninety (90) percent of the Contract Price of the Goods delivered to the Consignee shall be paid within 30 days of submission of following documents
- a. Three copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- b. Consignee Receipt Certificate in original issued by the authorized representative of the consignee;
- c. Manufacturer's/Supplier's warranty certificate & In-house inspection certificate.
- ii) On satisfactory installation, testing & commissioning: Ten (10) percent of the Contract Price of Goods received shall be paid within thirty (30) days of satisfactory installation & commissioning of the Goods and completion of training of the concerned personnel on operation and maintenance the equipment, upon submission of an invoice (indicating the SAMS as the Purchaser on behalf of FIND, the Contract number, credit number; description of payment and total amount, signed in original, stamped or sealed with the company stamp/seal) supported by the satisfactory installation, training & commissioning Certificate (Final Acceptance Certificate) issued by the Consignee.

### 32. Liquidated damages

- i. For delays- Subject to, other clauses, if the supplier fails to deliver any or all of the goods or perform the services within the time period(s) specified in the contract, the purchaser shall, without prejudice to its other remedies under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5 percent of the delivered price of the delayed goods unperformed services for each week of delay or part of thereof until actual delivery or performance up to a maximum deduction of 10% (ten) percent of the delayed goods or services contract price. However, the purchaser may consider termination of the contract or emergency purchases once the time specified is over.
- ii. For shortfall in equipment performance- Recovery shall be made from the supplier for not meeting the guaranteed performance/ productivity of the equipment during actual testing as detailed in "Technical Specifications". A sum equivalent to one percent of the cost of the equipment for each unit of shortfall in the guaranteed performance productivity where applicable.

#### 33. Return of rejected items of suppliers

The firm will take back the rejected items/ equipment from the consignees place and replace with fresh stock duly inspected within a reasonable period stipulated in the notice issued to the firm at their own cost up to the consignees or depots upon payment of testing charges, etc. The rejected stocks will be handed over by the respective consignees to the firm and payment of cost of the rejected stocks by the firm to the consignee. Failure of the firm may amount to the breach of contract on the part of the firm & firm shall be liable for recovery of consequent damages besides other remedies and rights available to the purchaser.

## 34. Penalty clause

If the firm fails to deliver any or all of goods or perform the services within the time period specified in the contract, the purchaser shall without prejudice to its other remedies under the contract, deduct from the contract price, as liquidated damages to 0.5 percent of the delivered price of the delayed goods or unperformed services for each week or delay or part thereof until actual delivery or performance, up to a maximum deduction of 10% (ten percent) of the delayed goods or services contract price. The purchaser may consider termination of the contract seeing its urgency.

#### 35. Taxes & Duties

Supplier shall be entirely responsible for all taxes, duties, fees, levies, entry tax, local taxes etc. incurred until delivery of the contracted goods till the final place of destination.

**36. VAT/PAN**: Copy of number allotted should also be submitted.

For Strategic Alliance Management Services Pvt. Ltd.

(Sanjay Rastogi)
Associate Director (MCS)

#### Annexures:

- I. Technical Specifications
- II. Price Form
- III. Schedule of Requirement and Consignee-wise Distribution
- IV. Manufacturer's Authorization Form
- V. Performance Security Form (Bank Guarantee)

# **Technical Specification of PCR Workstation, UV Light**

| Tender specifications  Description of function and use: This workstation is used to prepare reagents under clean conditions to avoid contamination with DNA.  Main specifications: Exterior dimensions (H X W X D) approx.: 700 x 750 x 800 mm Interior working area (W x D) approx.: 700 x 500 mm Interior working area (W x D) approx.: 700 x 500 mm Interior working area (W x D) approx.: 700 x 500 mm Interior made of stable formed stainless steel Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with  | NAME OF EQUIPMENT: PCR Workstation, UV Light                                  | Bidder's<br>Specification |
|--|---|---------------------------|
| Description of function and use:  This workstation is used to prepare reagents under clean conditions to avoid contamination with DNA.  Main specifications:  Exterior dimensions (H X W X D) approx.: 700 x 750 x 600 mm  Interior working area (W X D) approx.: 700 x 500 mm  Exterior: Stainless steel or powder coated metal interior made of stable formed stainless steel  Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation  Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm  UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each  Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CC Certifical and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spa |   | MANUFACTURER:             |
| Description of function and use: This workstation is used to prepare reagents under clean conditions to avoid contamination with DNA.  Main specifications:  Exterior dimensions (H X W X D) approx.: 700 x 750 x 600 mm  Interior working area (W x D) approx.: 700 x 500 mm  Exterior. Stainless steel or powder coated metal Interior made of stable formed stainless steel Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 mm  UV protection - UV-protective film on glass Panels Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key look for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certificad and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be sup | Tender specifications   | Type / Model:             |
| This workstation is used to prepare reagents under clean conditions to avoid contamination with DNA.  Main specifications:  Exterior dimensions (H X W X D) approx.: 700 x 750 x 600 mm  Interior working area (W x D) approx.: 700 x 500 mm  Exterior: Stainless steel or powder coated metal interior made of stable formed stainless steel  Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation  Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 mm  UV protection - UV-protective film on glass  Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each  Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage; 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates: issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along wit |   | Country of Origin:        |
| Main specifications:   |   |                           |
| Exterior dimensions (H X W X D) approx.: 700 x 750 x 600 mm  Interior working area (W x D) approx.: 700 x 500 mm  Exterior: Stainless steel or powder coated metal  Interior made of stable formed stainless steel  Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation  Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm  UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates: The manufacturer must have a management system certificat or ISO 9001  Manufacturer's Certificate: The manufacturer must have a management system certificate in Should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  |   | ation with DNA.           |
| Interior working area (W x D) approx.: 700 x 500 mm  Exterior: Stainless steel or powder coated metal Interior made of stable formed stainless steel Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm UV protection - UV-protective film on glass Panels Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529 Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Exterior: Stainless steel or powder coated metal Interior made of stable formed stainless steel Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm UV protection - UV-protective film on glass Panels Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529 Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted  | Exterior dimensions (H X W X D) approx.: 700 x 750 x 600 mm                   |                           |
| Interior made of stable formed stainless steel Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm UV protection - UV-protective film on glass Panels Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529 Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted   | Interior working area (W x D) approx.: 700 x 500 mm                           |                           |
| Should be equipped with a high efficiency flow-type bactericidal UV cleaner-recirculation which provide constant decontamination inside the cabinet during operation  Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm  UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each  Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| cleaner-recirculation which provide constant decontamination inside the cabinet during operation  Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm  UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each  Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| during operation  Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm  UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment:  Accessory to be quoted  |   |                           |
| Side panels transparent (no acryl, safety glass preferred), able to absorb wavelengths below 400 nm UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| wavelengths below 400 nm UV protection - UV-protective film on glass Panels Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529 Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory. Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| UV protection - UV-protective film on glass Panels  Overhead UV light for DNA decontamination, at least two lamps, 25 W each  Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W  Timer and key lock for UV lamp, timer operates only when key lock is on  UV exposure controlled by 24h digital timer  Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment:  Accessory to be quoted  |   |                           |
| Panels Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529 Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory. Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Overhead UV light for DNA decontamination, at least two lamps, 25 W each Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529 Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| Separate switchable, UV/Air Sterilizing Circulation unit, UV lamp 25 W Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse Electric needs Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Timer and key lock for UV lamp, timer operates only when key lock is on UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux Automatic switch-off of irradiated UV light when door opened At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1 Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| UV exposure controlled by 24h digital timer Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| Front opening with three adjustable positions for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment:  Accessory to be quoted   |   |                           |
| for ease of access  Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment:  Accessory to be quoted  | , ,   |                           |
| Overhead white light, 15 W at least 800 lux  Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment:  Accessory to be quoted   |   |                           |
| Automatic switch-off of irradiated UV light when door opened  At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment:  Accessory to be quoted   |   |                           |
| At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| At least two plug outlets built into the chamber, AC 230 ± 10V, 50 Hz and 5A fuse  Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| Electric needs  Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Supply voltage: 230 ± 10%, AC, 50/60 Hz, plug type adapted to those in India  Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment:  Accessory to be quoted   |   |                           |
| Consumption: Mainly depending on the electrical equipment used inside the work station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| station, max. 1200 W  CE Certified and Conform to electrical safety IEC-60601-1 and/or IEC 61010, UL 61010-1, EN 61010-1  Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  | station, max. 1200 W  |                           |
| Preferably, protection class in accordance with EN 60529  Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories:Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Quality and Safety standards met by the product offered have to be listed and copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001  One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| copies of the certificates issued by the authorized agencies to be enclosed for verification  Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| Manufacturer's Certificate: The manufacturer must have a management system certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  | , · · · · · · · · · · · · · · · · · · ·                                       |                           |
| certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   | verification  |                           |
| certified to ISO 9001 One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   | Manufacturer's Certificate: The manufacturer must have a management system    |                           |
| One certificate should state that the PCR Workstation has been calibrated at the factory.  Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted   |   |                           |
| Standard Maintenance Tools: All standard maintenance tools should be provided with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  |   |                           |
| with each unit  Spare parts/ Accessories: Spare Parts to be supplied along with equipment:  Accessory to be quoted   | factory.  |                           |
| Spare parts/ Accessories: Spare Parts to be supplied along with equipment: Accessory to be quoted  | Standard Maintenance Tools: All standard maintenance tools should be provided |                           |
| Accessory to be quoted   |   |                           |
|  |   |                           |
|  | Warranty: 3 Years warranty starts with installation on site                   |                           |

## Technical Specification of Desktop Deep freeze storage box with lid for 0.2ml PCR tubes

| Desktop Deep freeze storage box with lid | for |
|--|-----|
| 0.2ml PCR tubes                          |     |

Deep freeze storage box with lid for 0.2ml PCR tubes, autoclavable PP. Suitable for storing 0.2 ml single vessels (minimum 96 PCR tubes)/ 8-strips (12 tubes) / 12-strips (8 tubes) / 96-well PCR plates. Tube wells are easily identified by imprinted numbers and letters, removable, hinged lid, in different colours.

## **Price Form**

[The Bidder shall fill in this Price Form in accordance with the instructions given below]

| Sch. | BIDDER'S PRICES FOR GOODS (Price to be entered by Bidder):         |      |             |                       |                                    |                |  |
|------|--|------|-------------|-----------------------|------------------------------------|----------------|--|
| No.  |  |      |             |                       |                                    | INR            | •  |
|      | DESCRIPTION  | Unit | QTY.<br>(a) | UNIT<br>PRICE,<br>(b) | TOTAL<br>PRICE,<br>(c) = (a) x (b) | VAT/CST<br>(d) | TOTAL PRICE, FINAL Place of Destination* (e)=(c) + (d) |
|      | PCR-Workstation, UV light  | Nos. | 6           |                       |                                    |                |  |
| I    | Desktop Deep freeze<br>storage box with lid<br>for 0.2ml PCR tubes | Nos. | 70          |                       |                                    |                |  |

<sup>\*</sup>Total Price, till Final Place of Destination for quoted items should be inclusive of all applicable taxes and duties (Excise Duty, Custom Duty, Entry tax, Road Tax, Octroi Duty and any other local taxes etc.), only VAT/ CST to be quoted separately, as applicable. It should also include that all goods shall be duly insured up to the final destination.

## **Annexure-III**

## **Schedule of Requirement and Consignee Distribution List**

| SI.<br>No. | State             | Consignee details  | PCR-<br>Workstation,<br>UV light<br>(Qty.) | Desktop Deep<br>freeze storage<br>box with lid<br>for 0.2ml PCR<br>tubes (Qty.) |
|------------|-------------------|--|--|---|
| 1          | Karnataka         | Dr. B V Peerapur, Head, Dept of Microbiology, Raichur Institute of Medical Sciences, (RIMS) Hyderabad Road, RAICHUR Ph number: 09448139483 Email:peerapur_2003@yahoo.co.in   | 2  | 20  |
| 2          | Madhya<br>Pradesh | Dr. (Prof.) Shashi Gandhi Head of Department of Microbiology Gajra Raja Medical College, Opp. JAH Campus, Near KatoraTaal, Theme Road, Lashkar, Gwalior- 474009 Madhya Pradesh Office No: 07512332661 Ph number: 9826394482 Email ID: drshashigandhi@gmail.com&microbiology_department84@yahoo.com | 2  | 20  |
| 3          | Rajasthan         | Dr. PK Khatri Professor & Head of Microbiology Department, LPA Lab. IDI Block (Part of Department of Microbiology, SNMC, Jodhpur) Kamla Nehru Chest Hospital, Pal link road, Jodhpur- 342003 M :9001837262 drpkkhatri@yahoo.co.in&irlrjjdp@rntcp.org   | 0  | 10  |
| 4          | Tamil Nadu        | Dr. Jagatheeshwari (Director and HOD Microbiology Dept of Microbiology) Madurai Medical College Near Anna Bus stop Panagal Park Road Madurai: 625 020., Tamil Nadu. M: 9443408749 Dr. Surya Kumar Assistant Professor: 91-9894010207 mdumcmicro@gmail.com  | 2  | 20  |
|            |                   | Total Quantity   | 6  | 70  |

(Name of manufacturers)

## **MANUFACTURERS' AUTHORISATION FORM**

To,

| M/s Strategic Alliance Management Services Pvt. Ltd. (SAMS), 1/1 B, ChoudharyHetram House, Bharat Nagar, New Friends Colony, New Delhi 110025, INDIA                       |
|--|
| Dear Sir,  |
| RFQ No./Tender No.   |
| Wewho are established and reputable manufacturers of Having factories atand  |
| hereby, authorize M/s  |
| (Name and address of agents) to bid, negotiate and conclude the contract with you against IFB No for the above goods manufactured by us.                                   |
| No company or firm or individual other than M/s  |
| Are authorized to bid, negotiate and conclude the contract in regard to this business against this specific IFB as also for all business in the entire territory of India. |
| An agency commission of% included in the gross ex-work price is payable to M/s   |
| We hereby extend our full guarantee and warranty as per terms & conditions of contract for the goods offered for supply against this invitation for bid by the above firm. |
| Our other responsibilities include:  |
| (i)  |
| (ii)(Specify here-in detail manufacturer's responsibilities)   |
| The services to be rendered by M/s are as under:   |
| (i)  |
| (ii)(here specify the services to be rendered by the agent)  |
| Yours' faithfully,   |
| for and on behalf of M/s (Name)  |

Note: This letter of authorization should be on the letter head of the manufacturing concern and should be signed by a person competent and having the power of attorney to bind the manufacturer.

## PERFORMANCE SECURITY FORM (BANK GUARANTEE)

To The Strategic Alliance Management Services Pvt. Ltd. (SAMS), 1/1 B, Choudhary Hetram House, Bharat Nagar, New Friends Colony, New Delhi 110025, INDIA WHEREAS..... (Name and address of the supplier) (thereinafter called "the supplier") has undertaken, in pursuance of contract No...... dated ...... to supply (description of goods and services) (hereinafter called "the contract). AND WHEREAS it has been stipulated by you in the said contract that the supplier shall furnish you with a bank guarantee by a recognized bank for the sum specified therein as security for compliance with its obligations in accordance with the contract; AND WHEREAS we have agreed to give the supplier such a bank guarantee; NOW THEREFORE we hereby affirm that we are guarantors and responsible to you, on behalf of the supplier, up to a total of ..... ...... (amount of guarantee in words and figures). Such sum being payable in the types and proportions of currencies in which the contract price is payable, and we undertake to pay you, upon your First written demand declaring the supplier to be in default under the contract and without civil or argument, any sum or sums within the limits of (amount of guarantee) aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein. We hereby waive the necessity of your demanding the said debt from the supplier before presenting us with the demand. We further agree that no change or addition to or other modification of the terms of the contract to be performed there-under or of any of the contract documents which may be made between you and the supplier shall in any way release us from any liability under this guarantee and we hereby waive notice of any such change, addition or modification. This guarantee shall be valid until the date of issue of the "Acceptance Certificate" issued by the purchaser's representative. Signature and seal of the guarantor Place:

Date: