

Amendment No. 2 to ITB No.: SAMS - FIND- EQUIPMENT – PROCT - ICB - 03/2014 for International Competitive Bidding (ICB) for the Supply, Installation, Testing and Commissioning of Equipment for TB Laboratories across India

Following amendment is carried out in the ITB No.: SAMS - FIND- EQUIPMENT – PROCT - ICB - 03/2014

Sl. No.	For	Read
1	Schedule V: Autoclaves (Horizontal)	
	As per your specs vessel should be as per ASME, EN 13445 and EN 12953 for steam boiler with class B sterilization, whereas it should be ASME & EN13445 or EN 12953. ASME is must.	Vessel should be as per ASME, EN-13060-2 / EN-13445:2009 and EN 12953 for steam boiler with class B sterilization, implying that in addition to compliance either EN13445 or EN 12953, compliance to ASME is must.
2	Schedule VI: Autoclaves (Vertical)	
	As per your specs vessel should be as per ASME, EN 13060-2 / EN-13445:2009 and EN 12953, whereas it should be ASME, EN 13060-2 or EN 13445:2009 and EN 12953. ASME is must.	Vessel should be as per ASME, EN-13060-2 / EN-13445:2009 and EN 12953 for steam boiler with class B sterilization, implying that in addition to compliance either EN13445 or EN 12953, compliance to ASME is must.